

Certificate

QUMA

Composition of Calibration Standard QS9

order number Q100.103_QS9

ID no.: XXXX

	Name	CAS Nr
1	Ethylacetat	141-78-6
2	Isopropyl Acetat	108-21-4
3	Ethanol	64-17-5
4	N-Propyl Acetat	109-60-4
5	Propanol	71-23-8
6	1-Ethoxy-2-propanol	1569-02-4
7	IPA	67-63-0

Je / each 14,29 wt.% \pm 0,01%

Gesamtdichte / total Density d= 0,848

**36,34 mg/ components, when calibration with 3 μ l at 100 cm²
(72,68 mg/components by 3 μ l and 50 cm²).**

Qualität der Lösemittel besser 99,9 % garantiert
Quality of solvents better 99,9 % guaranteed

Haltbar bis / best before: XXXX

Hersteller/ Producer: QUMA Elektronik & Analytik GmbH / Germany

Prüfer: XXXX
Wuppertal / Germany XXXXX

